



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Caribbean Cement Company Limited

Sir Florizel Glasspole Boulevard
Rockfort, Kingston, Jamaica

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

TESTING

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

L2342

Certificate Number



ANAB Approval

Certificate Valid Through: 12/10/2021
Version No. 003 Issued: 11/06/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



ANSI National Accreditation Board

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Caribbean Cement Company Limited

Sir Florizel Glasspole Boulevard
Rockfort, Kingston, Jamaica

Andrew Stephenson astephenson@caribcement.com

TESTING

Valid to: **December 10, 2021**

Certificate Number: **L2342**

Construction Materials

| Specific Tests and/or Properties Measured | Specification, Standard, Method, or Test Technique | Items, Materials or Product Tested | Key Equipment or Technology |
|---|--|------------------------------------|---------------------------------|
| Compressive Strength | ASTM C109 | Cement | |
| Time of Setting by Vicat Needle | ASTM C191 | Cement | |
| Normal Consistency | ASTM C187 | Cement | |
| Air Content | ASTM C185 | Cement | |
| Autoclave Expansion - Soundness | ASTM C151 | Cement | |
| Flow | ASTM C1437 | Cement | |
| Fineness by Air Permeability Apparatus | ASTM C204 | Cement | |
| Chemical Analysis | ASTM C114 | Cement | X-Ray Fluorescence Spectrometry |
| Chemical Analysis – Loss on Ignition | ASTM C114 | Cement | |

Note:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. L2342.



Vice President